

Minutes of Meeting

Establishment of Hepatitis Clinics and GI Departments in Tertiary care hospitals all over the Punjab

1. Grievance committee meeting was held on 14-06-2018, under the chairmanship of Dr. Ghias Ul Hassan, Assistant Prof. of Gastroenterology to resolve the grievances submitted by the firms, for the purchase of medical equipment through ADP Scheme Establishment of Hepatitis Clinics and GI Departments in Tertiary care hospitals all over the Punjab.

2. Following committee members attended the meeting:

Dr. Bilal Nasir, Assistant Prof. of Gastroenterology

Dr. Asif Gul, Assistant Prof. of Gastroenterology

Dr. Qaiser Parveen, Retired AMS, Private Member

Dr. Suhail Qureshi, Manager BME, PMO

Mr. Awais Bin Tariq, EC/ Biomedical Engineer

3. The following items were discussed:-

Sr #	Item/ Package	Name of Firm	Remarks	Decision
1	Vital Sign Monitor	M/S Medi Urge	M/S Medi Urge submitted grievance against M/S Noor International that their quoted model of Vital Sign monitor SVM-7521 does not fulfil advertised technical specifications as it does not have Six Wave forms minimum. Committee discussed the case and found that as per PVMS and tender advertisement, six waveforms are required whereas the quoted model of Noor International have only four waveforms, therefore the committee accepted the grievance of the firm and rejected the quoted model of M/S Noor International.	Grievance of firm accepted. Quoted Model of M/S Noor International rejected.

Sr #	Item/ Package	Name of Firm	Remarks	Decision
2	Syringe & Infusion Pumps Package	M/S Iqbal & Co.	<p>M/S Iqbal & Company submitted grievance against M/S B Braun that:</p> <ol style="list-style-type: none"> 1. They have quoted old model of Syringe pump i.e. Perfusor Compact & Infusion Pump i.e Infusomat P, where as their new model is available which they are quoting in other hospital by the name of "Perfusor Compact Plus" & "Infusomat Compact Plus". 2. They have only quoted CIF rates and have not quoted FOR DDP rates. <p>Committee discussed the case and found that:</p> <ol style="list-style-type: none"> 1. M/S B Braun has quoted Perfusor Compact Plus & Infusomat P in their bid, which comply with advertised tender specifications and are technically approved. 2. In tender advertisement only CIF rates were required to be quoted for the said item. <p>Therefore the grievance of the firm is not accepted.</p>	Grievance not accepted.
			<p>M/S Iqbal & Company submitted grievance against M/S Medi Urge that they have only quoted CIF rates and have not quoted FOR DDP rates.</p> <p>Committee discussed the case and found that in tender advertisement only CIF rates were required to be quoted for the said item.</p> <p>Therefore the grievance of the firm is not accepted.</p>	Grievance not accepted.
3	Anti-Decubitus Ulcer Prevention mattress	M/S KASBN International	<p>M/S KASBN International submitted grievance against M/S Orblynx that their quoted Anti-decubitus Ulcer prevention Mattress does not fulfil advertised technical specifications due to lack of following:</p> <ol style="list-style-type: none"> 1. Four Header cells always run in a static way 2. CPR valve fitted in upper mattress area allows it to deflate the mattress in 10 seconds 3. Should have 13 firmness levels 4. Each cell is divided in two, with 7 cm in the lower section that acts always in static mode 5. Low Air loss. The mattress should have continuous air exit through 6. Pressure range 10-50mmHg or better 7. Patient weight capacity 250kg or better 8. Noise level 21dB or less 9. Minimum air input 7 litres/minute <p>Committee discussed the case and found that as per PVMS specifications, the quoted model of M/S Orblynx has only 02 firmness levels instead of 13 firmness levels and Pressure range is 32-42mmHg instead of 10-50mmHg. Whereas, all other points were verified from their technical documents.</p> <p>Hence the committee decided that the quoted model of M/S Orblynx failed to comply with advertised tender specifications and accepted grievance of the firm and rejected the quoted model of M/S Orblynx.</p>	Grievance of firm accepted. Quoted Model of M/S Orblynx rejected.

Sr #	Item/ Package	Name of Firm	Remarks	Decision
4	Pulse Oximeter	M/S Biotech Services	<p>M/S Biotech Services submitted grievance against M/S Delta Plus that their quoted Pulse Oximeter model PO-100B is not as per requirement, since as per specifications:</p> <ol style="list-style-type: none"> "Pulse rate measurement from 20-250 bpm" is required but their range starts from 25-300 bpm. Their quoted model is not manufactured by Advanced Instrumentations and this unit is actually manufactured by EDAN China with model name H100B. As per bidding document General Conditions of Contract "The medical equipment of USA must comply with FDA 510k". They do not have any Govt. reference and satisfactory report. M/S Delta Plus is not sole authorized distributor for Advanced Instruments. <p>Committee discussed the case and found that pulse rate measurement of 25-300 bpm is not as per tender specifications and rejected model of M/S Delta Plus on technical grounds.</p> <p>Committee called representative of M/S Delta Plus and asked them to clarify whether their quoted model was manufactured by Advanced Instrumentations, USA or EDAN, China and whether it had valid FDA 510K for quoted model, but the representative of the firm could not provide sufficient documentary evidence for either. M/S Delta Plus were able to demonstrate their government references and sole authorization of Advanced Instrumentation, USA.</p> <p>Hence the committee decided that the quoted model of M/S Delta Plus failed to comply with advertised tender specifications and accepted grievance of the firm and rejected the quoted model of M/S Delta Plus.</p>	Grievance of firm accepted. Quoted Model of M/S Delta Plus rejected.

Dr. Ghias ul Hassan, Assistant Prof. of Gastroenterology

Dr. Bilal Nasir, Assistant Prof. of Gastroenterology

Dr. Asif Gul, Assistant Prof. of Gastroenterology

Dr. Qaiser Parveen, Retd AMS, Private Member

Dr. Suhail Qureshi, Manager BME, PMO

Mr. Awais Bin Tariq, EC/ Biomedical Engineer