

No. 141 /LGH
To

dated Lahore the 20-4 2018

The Secretary,
Govt. of the Punjab,
Specialized Healthcare &
Medical Education, Lahore

Attention: (Focal Person) PPRA Website,
Govt. of the Punjab,
Specialized Healthcare &
Medical Education, Lahore

Subject: **MINUTES OF PRE BID COMMITTEE MEETING.**

With reference to the subject cited above.

It is therefore kindly requested to issue instruction to the focal person of PPRA Website to upload the minutes of Pre-Bid Committee held on 19-03-2018 at official Website of Health Department, Government of the Punjab, Lahore/PPRA for the Medical equipment's and others items through ADP scheme Establishment of Hepatitis Clinics and GI departments in all Tertiary care Hospitals in Punjab (Lahore General Hospital, Lahore).
(Minutes attached).




Prof. Dr. Ghias un Nabi Tayyab,
Project Director, Establishment of Hepatitis
Clinics and GI departments in all Tertiary care
Hospitals in Punjab (Lahore General Hospital,
Lahore)

No. _____ /LGH dated Lahore the _____ 2018

A copy is forwarded for information to the:-

1. Procurement Specialist, SHC & ME Dept., Govt. of the Punjab, Lahore.
2. Chief Planning Officer, SHC & ME Dept., Govt. of the Punjab, Lahore.



Prof. Dr. Ghias un Nabi Tayyab,
Project Director, Establishment of Hepatitis
Clinics and GI departments in all Tertiary care
Hospitals in Punjab (Lahore General Hospital,
Lahore)

Establishment of Hepatitis Clinics and GI departments in all Tertiary care Hospitals in Punjab

Minutes of Pre bid meeting

1. Pre Bid Committee meeting was held on 19-03-2018 under the chairmanship of Prof. Dr. Ghias Un Nabi Tayyab, Project Director, Establishment of Hepatitis Clinics and GI Sections all tertiary care hospitals of Punjab, regarding the Pre bid for the preparation of specifications for medical equipment and other items through ADP Scheme Establishment of Hepatitis Clinics and GI Sections all tertiary care hospitals of Punjab.

2. Following members attended the meeting:

Prof. Dr. Nauman Gillani, GMC, Gujranwala	Member
Prof. Dr. Kashif Malik, Sheik Hospital, FPGMI, Lahore	Member
Mr. Tayyab Farid, Procurement Specialist, SHC & ME Department	Member
Dr. M Junaid Mirza, PMO Community Medicine, PGMI	Member
Dr. Faisal Javed, DMS (ADP Purchase), LGH	Member
Mr. Usman Aziz Tarar, General Manager Procurement & Contracts, SHC & ME Department	Member
Mr. Syed Ali Asim, General Manager, Biomedical Engineering, SHC & ME Department	Facilitator
Mr. Zia Ul Haq, Technical Officer, Health Department	Facilitator
Mr. Natiq Ali Mehdi Zaidi, BME, LGH	Facilitator

3. With reference to meeting chaired by Honorable Secretary Health, SHC & ME Department, Govt. of the Punjab, on 12-03-2018, where it was decided that in the interest of fairness and open competition proprietary items like Spy Glass & Fibroscan be tendered openly.

4. Minutes of prequalification meeting held on 01-03-2018 as well as list of prequalified firms was uploaded on LGH and Health Department, Punjab websites and all participating firms were informed via email regarding pre bid meeting and last date for submission of bids.

5. Committee opened quotations of Vehicle Manufactureres and decided to send case to Austerity committee for approval.

6. The following items were discussed:-

Item/ Package No.	Name of Item	Qty	Comments
Medical Equipment			
Endoscopy Package			
1	Gastrointestinal video scopes	44	Committee decided that Water Jet function should be added to specifications, to be sent for standardisation.
	High definition colonovideoscope	22	As per PVMS
	Duodenovideo scope	23	As per PVMS
	Gastrointestinal video scopes(therapeutic)	15	Committee decided to change insertion tube diameter to "11.6 or less", and that working length should be "1030mm or more", to be sent for standardisation.
	Video system centre	20	As per PVMS
	Light source	20	As per PVMS
	Work station (sliding key board tray, LCD monitor arm, GI scope pole kit) Imported	19	As per PVMS
	Medical Grade Monitor (21" - 26")	19	Committee decided that monitor should be Trolley mounted, to be sent for standardisation



Item/ Package No.	Name of Item	Qty	Comments
Central Medical Gases Package			
2	Central medical gases/oxygen supply	17	Committee decided to get sites evaluated and make complete package specs based on each site's need.
	Bed Head Unit	170	
	Oxygen flow meter with humidifer bottle	680	
	Air venturisytem	680	
	Oxygen gas Cylinder with flow meter	170	
Sucker Machine Package			
3	Sucker machine Heavy Duty	17	As per PVMS
	Sucker Machine Light Duty	102	As per PVMS
Syringe & Infusion Pump Package			
4	Syringe Pump	340	As per PVMS. Committee decided to specify optional accessory "2 or more unit docking station" from PVMS 01 for each bed for a total quantity of 170 pieces, whereas optional TCI is not required.
	Infusion Pump	408	As per PVMS
Ultrasonic Cleaner/ Washer Disinfector Package			
5	ULTRASONIC CLEANER	14	Committee decided that since PVMS for this item does not exist, item name "Ultrasonic Cleaner for Flexible Endoscopy Accessories" and specifications with "6.9L or above capacity", to be sent for standardisation.
	Endoscopic washer and disinfector IMPORTED	25	As per PVMS. Committee decided to specify "Number of reprocessed scopes 02 at a time" & "Number of washing basins 01".
Radiology Package			
6	Image intensifier with video capturing capacity, lead paneling along with 6 lead aprons, thyroid shield and protective goggles	15	Committee decided to split this package to be bid separately, and since PVMS for this item does not exist, to change item name and specifications to "C-Arm Flat Panel Detector for Gastroenterology Procedures", to be sent for standardisation.
	Portable Ultrasound Machine	17	Committee decided to split this package to be bid separately, and to change item name and specifications to "Ultrasound Color Doppler for Gastroenterology", to be sent for standardisation.

Chaitan Singh

A.

W

J.

Minim

Alley

km

g

Item/ Package No.	Name of Item	Qty	Comments
Beds & Trolleys package			
7	ICU/HDU Electric Bed with Bedside Cabinet	170	Committee decided that since PVMS for this does not exist, specifications to be sent for standardisation.
	Stretcher Trolley/ Patient Trolley Stretcher/ Emergency Patient Shifting Trolley/ Hi Low Patient Trolley	476	As per PVMS
	Patient Bed , IV stand, Bed locker	680	Committee decided to split this package to be bid separately, so that Imported beds may be quoted separately from US/EU/JPN items. Specs to be sent for standardisation.
8	Electro surgical unit with APC USA/ EU/ Japan	14	Committee decided that since PVMS for this does not exist, item name to be changed and specifications to be sent for standardisation of "Argon Plasma with Electrosurgical Unit for Gastrointestinal Endoscopy".
9	Vital Sign Monitor	289	As per PVMS. Committee decided on 12" Non-invasive Monitor with 5 parameter.
10	ECG Machine 12 channel	51	As per PVMS
11	Ultra Sonic Nebulizer	119	As per PVMS
12	Mobile Resuscitation Trolley with Defibrillator	34	As per PVMS
13	Transport Trolley with ventilator and monitor	17	As per PVMS
14	Pulse oximeter	340	As per PVMS
15	PCR machine USA/ EU/ Japan	15	Committed opted for following PKLI specs for Fully Automated System with programmable channels PCR, with the provision of B, C, D testing , with FDA approved kits. Committee decided to get comment from Health Department to shift procurement of Pathology items to Reagent-based basis.
16	Blood gas Analyzer	17	Committee decided to upgrade the specifications to Blood Gas Analyzer with Electrolytes. Committee decided to get comment from Health Department to shift procurement of Pathology items to Reagent-based basis.
17	Elisa assays	11	Name should be changed from assay to analyzer. Committee decided to get comment from Health Department to shift procurement of Pathology items to Reagent-based basis
Cost of Other Store package			
18	Stethoscope (Locally provided)	68	Specs as per standard
	Standard Ophthalmoscope	34	
	BP apparatus with stand	221	
	Laryngoscope set	170	
	Percussion Hammer	68	

S. S. S. S.

J.

J.

Thur

Chian S. S.

Chian S. S.

Item/ Package No.	Name of Item	Qty	Comments
	Glucometer	51	Committee decided that item is superfluous and readily available on strip- basis in every hospital and should be dropped from PC1.
19	Ripple mattress	340	Committee decided that since PVMS for this does not exist, item name to be changed and specifications to be sent for standardisation of "Anti-Decubitus Ulcer Prevention Mattress" of US/EU/JPN origin
20	Stethoscope USA/ EU/ Japan	153	Committee decided that Stethoscope should be Littman or ADC brand

Hospital Furniture Package

21	Patient attendant bench	680	Specs as per standard
	Patient Examination Couch	51	
	X-Ray Illuminator 4 films	68	
	Steel almairah	68	
	Patient Screen Partition Folding	34	
	Medicine Trolley	51	
	Wheel Chairs	68	
	IV Stand	102	

Office Furniture Package

22	Central Table set special	17	Specs as per standard
	Computer Table	168	
	Office Table	206	
	Executive Office Table with side rack	68	
	Executive Chairs	136	
	Office Chairs	850	
	Office Chair Foam Seated	8	
	Computer Chair	168	
	Class room chair	850	
	Revolving Chair	70	
	Revolving Stool	68	
	Patient waiting chairs	1020	
	2 seater sofa	17	
	3 seater sofa	17	
	Racks & Shelves	170	
	Shoe Racks	68	

S. st
after

rch

J.

Jr.

thw — in

Chen Sun

Alan

km

Dr

Item/ Package No.	Name of Item	Qty	Comments
	Notice board	51	
Plant & Machinery package			
23	Electric Water Cooler 60 Gallon	17	Specs as per standard
	Wet & Dry Industrial Heavy Duty Vacuum Cleaner	34	
	Refrigerator 11 Cubic	34	
	Water dispenser	34	
General Store Items package			
24	LCD 42"	17	Specs as per standard
	Moveable dustbin	34	
	Dust Bin Paddle Led	87	
	Dustbin	1156	
	Flower Pot	170	
	Torch 3-cell	68	

Prof. Dr. Ghias un Nabi Tayyab, Principal, AMC/ PGMI/ LGH

Prof. Dr. Nauman Gillani, GMC, Gujranwala

Prof. Dr. Kashif Malik, Sheik Hospital, FPGMI, Lahore

Mr. Tayyab Farid, Procurement Specialist, SHC & ME Department

Dr. M Junaid Mirza, PMO Community Medicine, PGMI

Dr. Faisal Javed, DMS (ADP Purchase)

Mr. Usman Aziz Tarar, General Manager Procurement & Contracts, SHC & ME Department

Mr. Syed Ali Asim, General Manager, Biomedical Engineering, SHC & ME Department

Mr. Zia Ul Haq, Technical Officer, Health Department

Mr. Natiq Ali Mehdi Zaidi, BME, LGH